



ELTEC INSTRUMENTS, INC

ELTECdata #116

Notes On Apparent Temporal Baseline Shifts From Pyroelectric Detectors

First, the detector is a finite mass composed of materials with different specific heats and different thermal conductivities. When a pyroelectric detector attains thermal equilibrium, the output of the detector is "zero". The equilibrium is dynamic in the sense that the detector base, can and filter are always radiating infrared to the sensing crystal which is itself radiating. Thus assuming no conductive/convective heat transfer, the net infrared radiation exchange is also zero.

A change in the thermal environment of the detector will naturally change the temperature of the detector. If the rate of change is well below the thermal and electrical time constants of the detector, no signal will be produced. (The developed charge on the electrodes will leak to ground as it is produced.)

It is often said that a pyroelectric detector produces a signal only if it experiences a change. It is well to note that a rapid and steady warming (or cooling) of the detector can be equated to a series of step changes in one direction. If there were an identical detector with the sensing crystal "blinded" to external infrared radiation, it would be responding only to the ambient temperature change and its signal could be inverted and subtracted from the output of the first detector for a combined output representative only of the pulse. (Actually, Eltec's Model 405, 407 and Model 446 Thermally Compensated Detectors operate much as described).

For uncompensated detectors receiving strong pulses of infrared radiation, the situation is more difficult. For a detector with the sensing crystal heat sunk to a metal base, an extended series of high energy pulses (all of equal energy), causes the output of the detector to behave as follows: The first pulse strikes and produces a specific output. Before the detector crystal and thermal mass can dissipate the energy absorbed, a second pulse strikes. An output is produced from the second pulse. The output from the second pulse represents the magnitude of the pulse plus a baseline increase. After a series of pulses ends, the crystal/thermal mass assembly achieves equilibrium and the initial baseline is restored. For a detector with the sensing crystal cantilevered or loop mounted the behavior is similar, but achieving the initial baseline restoration will be longer.

Dependencies/Qualifications: If the combination of pulse energy and repetition rate infuse energy to the crystal/mass assembly faster than it can be dissipated the assembly will increase in temperature until damage occurs. Also note that differing applications dictate different detector designs. In some detectors the crystals are elevated above the substrate by thin gold wires to thermally isolate the crystal. This greatly increases the thermal time constant of the crystal and brings a corresponding increase in responsivity to weak infrared radiation pulses. However, this thermally- isolated design will show greater baseline shifts than a laser detector with a heat-sunk crystal. Also note that the electrical termination of the leads from the electrodes on the crystal influence the transfer function and time dependencies (Crystal capacitance in conjunction with load resistance).

What is the detector user to do? First, be aware that a thermal detector involves thermal time constants, heat dissipation, etc. Second, after examining the specific application, choose a detector with the best thermal design for the application. Third, test the system changing, in turn, ambient temperature, pulse intensity, and repetition rate such that the actual parameters of the intended application are bracketed and appropriate normalization of the data can be made.

NOTICE: The information provided herein is believed to be reliable. However, ELTEC Instruments, Inc. assumes no responsibility for inaccuracies or omissions. ELTEC Instruments, Inc. assumes no responsibilities for the use of the information, and all use of such information shall be entirely at the User's own risk. Publication of this information does not imply any authority or license for its use nor for any infringement of patents or rights of others which may result from its use.



**ELTEC Instruments, Inc. P.O. Box 9610 Daytona Beach, Florida 32120-9610 U.S.A.
Tel (USA and Canada): (800) 874-7780 Tel (Outside USA): (386) 252-0411 Fax: (386) 258-3791
Web: www.eltecinstruments.com E-Mail: Sales@eltecinstruments.com**